

FEB 24 2011
3/17/11

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130141 ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name Lynne Gralewski-Baugus	ID# 6787 ✓ Date 20110214 ✓
A Agency 673rd Security Forces Squadron Phone # 907-552-4000	
Instrument Location Joint Base Elmendorf-Richardson	
B Alco S/N A090740 ✓ Target Value .083 ✓ High Pressure 500 psi	
Alco Test Values	.083 ✓ 1st Alco
	.083 ✓ 2nd Alco
Signature	<i>Lynne Gralewski-Baugus</i>
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 17th day of MAR, 2011.

Daniel L. Nicholas
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130141

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130141

FEBRUARY 14, 2011

OPERATOR'S NAME:

GRALEWSKI-BAUGUS/LYNNE/M

OPERATOR'S NUMBER: 6787

SUBJECT'S LAST NAME:

VERIFICATION OF CALIBRATION

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: JBR4

CASE/REPORT: VOC

TEST TYPE: U

ALCO TARGET VALUE: .083

ALCO S/N: A090740

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 29.93 in
ALCO TARGET .083 13:27
BLANK TEST .000 13:28
INTERNAL STANDARD VERIFIED 13:28
ALCO TV 29.93 in .083 13:29
BLANK TEST .000 13:29
SUBJECT SAMPLE .000 13:30
BLANK TEST .000 13:31
ALCO TV 29.93 in .083 13:31
BLANK TEST .000 13:32

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130141

FEBRUARY 14, 2011

TIME 13:15

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 41c
BAROMETER: 29.93 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLNMOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~